

Replacement of the cartridge in HydroFlex

With a few simple steps, you can replace the cartridge in HydroFlex hydrogen reference electrode.

For the functioning of the electrode it is important to stick to the following instructions:



Remove the cartridge.



Take the new cartridge and grease the thread.



Screw the cartridge back into the head of the electrode.



You can activate the electrode now.