

Manual

Temperature Control Box



Version 1 - 10.02.2018
Update on 15.02.2021

Gaskatel GmbH
Lilienthalstrasse 146
Building 11
34123 Kassel - Germany

+49 561 / 59 190

- 1 Heating of half cells 3
- 2 PTC Heating elements 3
 - 2.1 General remarks 3
 - 2.2 Used elements 3
 - 2.3 Parallel oder serial connection..... 4
- 3 Temperature-Measurement..... 4
- 4 Temperatur-Control 4
- 5 Power supply 5
- 6 Setting-up Operation 6
 - 6.1 Operating Conditions 6
 - 6.2 Warnings 6

1 Heating of half cells

The electrolyte in the equipment should be heatable.

Electrical heating is recommended, because otherwise you may contaminate your electrolyte with the heating medium (oil, water with surfactants).

Due to the electric conductivity of electrolytes you should only apply safety extra low voltages. A special protection is not needed under these circumstances.

One has to use self limiting heating elements not to exceed the maximum temperature of the plastic we use in our test equipment

For the PP half cell the elements are limited to 130°C

For the PTFE half cell the elements are limited to 240°C .

2 PTC Heating elements

2.1 General remarks

PTC heating elements are self limiting. Below the maximum temperature they have a low impedance. Above that temperature the impedances increases and the current reduces.


2.2 Used elements

The heating elements are delivered from DBK:

PP Test cell




- PP-heating element, custom made product Gaskatel, 12-24V
- Resistance of one element 2.2 Ohm



- PTC heating element DBK HP 06 2/10-24
- Resistance of one element 3 Ohm

PTFE Test cell



- PTC heating element DBK HP 06 2/22 12-24V
- Resistance of one element 5 Ohm

The heating elements are assembled with Wago connection terminals.

2.3 Parallel oder serial connection

You may connect the heating elements in parallel or serial.

At the PP test cell the elements are always connected in parallel.

At the PTFE test cell we recommend a serial connection up to temperatures of 110°C. If you are looking for higher temperatures you should change to a parallel connection.

3 Temperature-Measurement

The temperature sensor should have a protection cover out of PTFE. Other sensors with for example metallic protection covers should not be used because of the corrosion attack by the electrolyte.

We can recommend the following sensors:

- PT100 from Omega HSRTD 3-100-A-1m
- PT100 from TC Direkt: FEP Pt100

4 Temperatur-Control

Adjust the nominal temperature

Press the ▲▼ key to adjust the nominal temperature. With the ◀ key you can jump to the next decimal place

Auto-Tuning:

The PID control may adjust the optimal values on its own. Press in that case the AT key for 2 seconds.

Set-Up.

Press the Set key for 2 seconds. You will enter the set up manual. Additional informations are in the Sestos manual. The most important factors are:

CTrL: 1 On/Off; 2 Auto adjustment, 3 PID
oPH : Upper output limit

If the temperature overshoots that happens because of a too large power supply. You can adjust this power supply value with the parameter **oPH** (Output upper limit 0 bis 220). Reduce that value a run again the auto tuning by adjusting the CTrL value to „2“.

Tabelle 1: Typical PID Values for the PP half cell

		Temperature [°C]					
		30	40	50	60	70	80
M50	Integral („I“)	1500	300	71	40	25	166
P	Differential („P“)	920	1940	1	4844	9998	888
T	Hysteresis time („D“)	414	364	340	218	266	391
CTI	Control period	20	60	60	60	120	20
oPH	Output upper limit	10	40				

Tabelle 2: Typisal PID Values for the PTFE half cell (parallel connection)**

		Temperature [°C]					
		30	50	70	90	110	130 **
M50	Integral („I“)	750	214	181	150	133	185
P	Differential („P“)	882	626	624	734	950	788
T	Hysteresis time („D“)	618	1109	507	592	390	296
CTI	Control period	40	40	40	60	60	40
oPh	Output upper limit	10	30	40	100	100	80

The power controller is a solid state relay (A Sestos SSR 521). This reduces the current / voltage spikes which may interfere with the sensitive inputs of the potentiostat.

5 Power supply

Power supply 24V 240W integrated.

You should use banana plugs to connect the cell with the temperature control bos. The polarity is not important. The open ends of the cable are inserted into the Wago contact terminals of the cell.

At these contact terminals you can change from parallel to serial connection

6 Setting-up Operation

6.1 Operating Conditions



Nominal voltage

The maximum voltage should not exceed 24 V.



Nominal current

The maximum current is 25Ampere for the SSR relais.

6.2 Warnings



Hot surfaces

The cell heaters and / or the applied cell current can heat up the test cell far beyond 80°C.
Do not place heat sensitive things on the test cell.
Do not place inflammable things near the test cell.
Pay attention to a correct lining of your measurement and heating cable.
Select cable insulations with a reasonable temperature specification.
Else there is a risk of an electrical shortcut and fire hazard.

Laboratory use



The half cell is defined solely for use in a laboratory environment.
The laboratory environment must be conform to the safety data sheets and specifications of your electrolyte.

Protective equipment



The operator of the half cell must be dressed with adequate laboratory protective equipment according the safety data sheets and specifications of your electrolyte.
A pressure surge on the gas inlet or reach of boiling point may lead to electrolyte sputtering out of the half cell.
Electrolyte vapour according vapour pressure curve is being emitted permanently by the half cell.